

Bia Diagnostics 480 Hercules Drive Suite 101 Colchester, VT 05446

(802) 540-0148 https://www.biadiagnostics.com/ Lic# TLAB0029

QA Testing


1 of 1

SUZYS GIFT CBD HEMP PREROLL

Sample ID: BIA250207S0020 Strain: SUZYS GIFT

Matrix: Plant Type: Preroll Sample Size: 1.317 g Lot#: 00594_2022_02_PR01 Produced: Collected: Received: 02/07/2025 Completed: 02/13/2025 Batch#:

Client Vermont Botanical Farm LLC Lic.# , VT 05647



Cannabinoids

0.42% Total THC	14.46% Total CBD		17.28% Total Cannabinoids
Analyte LOQ	Mass	Mass	
CBDVa 0.0001 CBDV 0.0001 CBDa 0.0001 CBGa 0.0001 CBGa 0.0002 CBD 0.0002 CBN 0.0002 CBC 0.0002 THCa 0.0003 Total THC Total CBD Total CBD	% 0.29 0.13 10.89 0.23 0.11 4.91 <loq <loq 0.25 <loq <loq 0.27 0.20 0.42 14.46 17.28</loq </loq </loq </loq 	mg/g 2.9 1.3 108.9 2.3 1.1 49.1 <loq <loq <loq 2.5 <loq <loq 2.7 2.0 4.21 144.59 172.82</loq </loq </loq </loq </loq 	

Analyst: 056

Cannabinoids Methodology: High Performance Liquid Chromatography (HPLC) using PerkinElmer FLEXAR 🎟 with Photo Diode Array Detector (PDA)

Total CBD and total THC are calculated values, to account for assumed decarboxylation from the acid form (THCA or CBDA) to the neutral form, causing weight loss of the acid group. These values are calculated as follows:

TotalTHC=(THCAx0.877)+Δ9-THC

Total CBD = (CBDA x 0.877) + CBD Reagent Blanks: < LOQs for all analytes

LOQ = The lowest quantity that this method can reliably detect. Any cannabinoid that was not detected is assumed to be less than the stated LOQ (<LOQ).

All results reflect dry weight of material, based on % moisture of the sample. Measurement of Uncertainty (MU): the parameter, associated with the result of a measurement, that characterizes the dispersion of the values that could reasonably be attributed to the particular quantity subject to measurement. Δ9-THC MU = ±0.005% Total THC MU = ±0.007%

All other cannabinoid MU values are available upon request.

All moisture and water activity analysis is determined by dewpoint measurement using an AQUALAB water activity meter.



sample as received.

ulle Luke Emerson-Mason

Confident LIMS All Rights Reserved coa.support@confidentlims.com (866) 506-5866 www.confidentlims.com



02/13/2025 This report shall not be reproduced except in full without approval of the laboratory. This is to provide assurance that parts of a report are not taken out of context. All results apply to this

Laboratory Director

Completed